

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB073123\
 Data File : FB030227.D
 Signal(s) : FID2B.CH
 Acq On : 31 Jul 2023 13:37
 Operator : YP/AJ
 Sample : 03810-10MSD
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 31 15:35:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB072423.M
 Quant Title :
 QLast Update : Tue Jul 25 06:41:02 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.787	258138	21.002 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.702	411130	24.887 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	482413	25.997 ng/ml
3) t n-Heptane	7.745	163015	8.474 ng/ml
4) t Benzene	7.882	226308	9.620 ng/ml
6) t Toluene	10.616	656145	28.631 ng/ml
7) t Ethylbenzene	13.053	207640	9.674 ng/ml
8) t m-Xylene	13.185	404897	19.253 ng/ml
9) t O-Xylene	13.912	392133	19.107 ng/ml
10) t 1,2,4-Trimethylbenzene	16.188	359218	18.352 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB073123\
Data File : FB030227.D
Signal(s) : FID2B.CH
Acq On : 31 Jul 2023 13:37
Operator : YP/AJ
Sample : 03810-10MSD
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 31 15:35:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB072423.M
Quant Title :
QLast Update : Tue Jul 25 06:41:02 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

