

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080119\
 Data File : FB021891.D
 Signal(s) : FID2B.CH
 Acq On : 1 Aug 2019 13:07
 Operator : DD/AJ
 Sample : BSF0801S1
 Misc : 5.02/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 02 00:39:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB071519.M
 Quant Title :
 QLast Update : Mon Jul 15 13:30:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.989	269497	19.748 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.862	709444	25.727 ng/ml
2) t 2,2,4-Trimethylpentane	7.609	764033	28.903 ng/ml
3) t n-Heptane	7.949	231989	9.356 ng/ml
4) t Benzene	8.083	280269	9.593 ng/ml
6) t Toluene	10.830	594938	28.914 ng/ml
7) t Ethylbenzene	13.275	206136	9.908 ng/ml
8) t m-Xylene	13.408	442128	19.902 ng/ml
9) t O-Xylene	14.137	458321	19.925 ng/ml
10) t 1,2,4-Trimethylbenzene	16.416	478068	20.277 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080119\
Data File : FB021891.D
Signal(s) : FID2B.CH
Acq On : 1 Aug 2019 13:07
Operator : DD/AJ
Sample : BSF0801S1
Misc : 5.02/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 02 00:39:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB071519.M
Quant Title :
QLast Update : Mon Jul 15 13:30:07 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.862 min
Delta R.T.: -0.001 min
Response: 709444
Conc: 25.73 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.609 min
Delta R.T.: -0.004 min
Response: 764033
Conc: 28.90 ng/ml




#3 n-Heptane

R.T.: 7.949 min
Delta R.T.: -0.002 min
Response: 231989
Conc: 9.36 ng/ml




#4 Benzene

R.T.: 8.083 min
Delta R.T.: -0.003 min
Response: 280269
Conc: 9.59 ng/ml




#5 AAA-TFT

R.T.: 8.989 min
Delta R.T.: -0.003 min
Response: 269497
Conc: 19.75 ng/ml




#6 Toluene

R.T.: 10.830 min
Delta R.T.: -0.003 min
Response: 594938
Conc: 28.91 ng/ml




#7 Ethylbenzene

R.T.: 13.275 min
Delta R.T.: -0.003 min
Response: 206136
Conc: 9.91 ng/ml




#8 m-Xylene

R.T.: 13.408 min
Delta R.T.: -0.003 min
Response: 442128
Conc: 19.90 ng/ml



#9 O-Xylene

R.T.: 14.137 min
Delta R.T.: -0.004 min
Response: 458321
Conc: 19.92 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.416 min
Delta R.T.: -0.003 min
Response: 478068
Conc: 20.28 ng/ml