

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080320\
 Data File : FB024713.D
 Signal(s) : FID2B.CH
 Acq On : 3 Aug 2020 16:55
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 04 02:17:03 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
 Quant Title :
 QLast Update : Thu Jul 09 10:49:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.802	513420	22.288 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.690	961770	32.500 ng/ml
2) t 2,2,4-Trimethylpentane	7.419	1037814	33.150 ng/ml
3) t n-Heptane	7.761	343310	11.042 ng/ml
4) t Benzene	7.895	428464	11.241 ng/ml
6) t Toluene	10.633	1217247	33.291 ng/ml
7) t Ethylbenzene	13.073	374171	10.955 ng/ml
8) t m-Xylene	13.206	729248	21.609 ng/ml
9) t O-Xylene	13.934	720502	21.922 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	608906	22.573 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080320\
Data File : FB024713.D
Signal(s) : FID2B.CH
Acq On : 3 Aug 2020 16:55
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 04 02:17:03 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
Quant Title :
QLast Update : Thu Jul 09 10:49:06 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.690 min
Delta R.T.: 0.002 min
Response: 961770
Conc: 32.50 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.419 min
Delta R.T.: 0.003 min
Response: 1037814
Conc: 33.15 ng/ml




#3 n-Heptane

R.T.: 7.761 min
Delta R.T.: 0.005 min
Response: 343310
Conc: 11.04 ng/ml




#4 Benzene

R.T.: 7.895 min
Delta R.T.: 0.003 min
Response: 428464
Conc: 11.24 ng/ml




#5 AAA-TFT

R.T.: 8.802 min
Delta R.T.: 0.004 min
Response: 513420
Conc: 22.29 ng/ml




#6 Toluene

R.T.: 10.633 min
Delta R.T.: 0.004 min
Response: 1217247
Conc: 33.29 ng/ml




#7 Ethylbenzene

R.T.: 13.073 min
Delta R.T.: 0.004 min
Response: 374171
Conc: 10.95 ng/ml




#8 m-Xylene

R.T.: 13.206 min
Delta R.T.: 0.004 min
Response: 729248
Conc: 21.61 ng/ml



#9 O-Xylene

R.T.: 13.934 min
Delta R.T.: 0.005 min
Response: 720502
Conc: 21.92 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.005 min
Response: 608906
Conc: 22.57 ng/ml