

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080320\
 Data File : FB024716.D
 Signal(s) : FID2B.CH
 Acq On : 3 Aug 2020 18:28
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 04 02:17:35 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
 Quant Title :
 QLast Update : Thu Jul 09 10:49:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.801	465854	20.223 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	923875	31.219 ng/ml
2) t 2,2,4-Trimethylpentane	7.420	985207	31.469 ng/ml
3) t n-Heptane	7.761	326318	10.495 ng/ml
4) t Benzene	7.895	418663	10.984 ng/ml
6) t Toluene	10.633	1188437	32.503 ng/ml
7) t Ethylbenzene	13.073	360870	10.565 ng/ml
8) t m-Xylene	13.206	707523	20.966 ng/ml
9) t O-Xylene	13.933	698557	21.254 ng/ml
10) t 1,2,4-Trimethylbenzene	16.211	583467	21.630 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080320\
Data File : FB024716.D
Signal(s) : FID2B.CH
Acq On : 3 Aug 2020 18:28
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 04 02:17:35 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
Quant Title :
QLast Update : Thu Jul 09 10:49:06 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.691 min
Delta R.T.: 0.003 min
Response: 923875
Conc: 31.22 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.420 min
Delta R.T.: 0.004 min
Response: 985207
Conc: 31.47 ng/ml




#3 n-Heptane

R.T.: 7.761 min
Delta R.T.: 0.005 min
Response: 326318
Conc: 10.50 ng/ml




#4 Benzene

R.T.: 7.895 min
Delta R.T.: 0.003 min
Response: 418663
Conc: 10.98 ng/ml




#5 AAA-TFT

R.T.: 8.801 min
Delta R.T.: 0.003 min
Response: 465854
Conc: 20.22 ng/ml




#6 Toluene

R.T.: 10.633 min
Delta R.T.: 0.004 min
Response: 1188437
Conc: 32.50 ng/ml




#7 Ethylbenzene

R.T.: 13.073 min
Delta R.T.: 0.004 min
Response: 360870
Conc: 10.57 ng/ml




#8 m-Xylene

R.T.: 13.206 min
Delta R.T.: 0.004 min
Response: 707523
Conc: 20.97 ng/ml



#9 O-Xylene

R.T.: 13.933 min
Delta R.T.: 0.004 min
Response: 698557
Conc: 21.25 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.211 min
Delta R.T.: 0.004 min
Response: 583467
Conc: 21.63 ng/ml