

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080320\
 Data File : FB024707.D
 Signal(s) : FID2B.CH
 Acq On : 3 Aug 2020 13:27
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Aug 04 02:16:05 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
 Quant Title :
 QLast Update : Thu Jul 09 10:49:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.799	458554	19.907 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	900739	30.437 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	997499	31.862 ng/ml
3) t n-Heptane	7.757	324889	10.449 ng/ml
4) t Benzene	7.892	397208	10.421 ng/ml
6) t Toluene	10.630	1153709	31.553 ng/ml
7) t Ethylbenzene	13.070	354138	10.368 ng/ml
8) t m-Xylene	13.203	699118	20.717 ng/ml
9) t O-Xylene	13.931	690693	21.015 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	583106	21.617 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080320\
Data File : FB024707.D
Signal(s) : FID2B.CH
Acq On : 3 Aug 2020 13:27
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Aug 04 02:16:05 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
Quant Title :
QLast Update : Thu Jul 09 10:49:06 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 900739
Conc: 30.44 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.002 min
Response: 997499
Conc: 31.86 ng/ml




#3 n-Heptane

R.T.: 7.757 min
Delta R.T.: 0.001 min
Response: 324889
Conc: 10.45 ng/ml




#4 Benzene

R.T.: 7.892 min
Delta R.T.: 0.000 min
Response: 397208
Conc: 10.42 ng/ml




#5 AAA-TFT

R.T.: 8.799 min
Delta R.T.: 0.000 min
Response: 458554
Conc: 19.91 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 1153709
Conc: 31.55 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.001 min
Response: 354138
Conc: 10.37 ng/ml



#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: 0.001 min
Response: 699118
Conc: 20.72 ng/ml




#9 O-Xylene

R.T.: 13.931 min

Delta R.T.: 0.002 min

Response: 690693

Conc: 21.01 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min

Delta R.T.: 0.001 min

Response: 583106

Conc: 21.62 ng/ml