

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080620\
 Data File : FB024723.D
 Signal(s) : FID2B.CH
 Acq On : 6 Aug 2020 12:57
 Operator : DD/AJ
 Sample : L3553-03MS
 Misc : 5.04g/5.00ml DI WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Aug 07 02:26:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
 Quant Title :
 QLast Update : Thu Jul 09 10:49:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	535644	23.253 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	877710	29.659 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	864892	27.626 ng/ml
3) t n-Heptane	7.757	276067	8.879 ng/ml
4) t Benzene	7.891	418335	10.975 ng/ml
6) t Toluene	10.629	1136830	31.091 ng/ml
7) t Ethylbenzene	13.069	338704	9.916 ng/ml
8) t m-Xylene	13.202	661103	19.590 ng/ml
9) t O-Xylene	13.930	652974	19.867 ng/ml
10) t 1,2,4-Trimethylbenzene	16.207	525691	19.488 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080620\
Data File : FB024723.D
Signal(s) : FID2B.CH
Acq On : 6 Aug 2020 12:57
Operator : DD/AJ
Sample : L3553-03MS
Misc : 5.04g/5.00ml DI WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Aug 07 02:26:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
Quant Title :
QLast Update : Thu Jul 09 10:49:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 877710
Conc: 29.66 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: 0.000 min
Response: 864892
Conc: 27.63 ng/ml




#3 n-Heptane

R.T.: 7.757 min
Delta R.T.: 0.001 min
Response: 276067
Conc: 8.88 ng/ml



#4 Benzene


R.T.: 7.891 min
Delta R.T.: 0.000 min
Response: 418335
Conc: 10.97 ng/ml



#5 AAA-TFT


R.T.: 8.798 min
Delta R.T.: 0.000 min
Response: 535644
Conc: 23.25 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.629 min
Delta R.T.: 0.000 min
Response: 1136830
Conc: 31.09 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: 0.000 min
Response: 338704
Conc: 9.92 ng/ml



#8 m-Xylene


R.T.: 13.202 min
Delta R.T.: 0.000 min
Response: 661103
Conc: 19.59 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: 0.000 min
Response: 652974
Conc: 19.87 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.207 min
Delta R.T.: 0.000 min
Response: 525691
Conc: 19.49 ng/ml