

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080620\
 Data File : FB024726.D
 Signal(s) : FID2B.CH
 Acq On : 6 Aug 2020 15:13
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc : 5.07g/5.00ml DI WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Aug 07 02:26:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
 Quant Title :
 QLast Update : Thu Jul 09 10:49:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.799	492858	21.396 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.691	898201	30.352 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	1008335	32.208 ng/ml
3) t n-Heptane	7.758	333426	10.724 ng/ml
4) t Benzene	7.893	405263	10.632 ng/ml
6) t Toluene	10.631	1138780	31.145 ng/ml
7) t Ethylbenzene	13.070	345122	10.104 ng/ml
8) t m-Xylene	13.203	675996	20.031 ng/ml
9) t O-Xylene	13.931	664449	20.216 ng/ml
10) t 1,2,4-Trimethylbenzene	16.208	539763	20.010 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080620\
Data File : FB024726.D
Signal(s) : FID2B.CH
Acq On : 6 Aug 2020 15:13
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc : 5.07g/5.00ml DI WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Aug 07 02:26:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070920.M
Quant Title :
QLast Update : Thu Jul 09 10:49:06 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.691 min
Delta R.T.: 0.003 min
Response: 898201
Conc: 30.35 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.002 min
Response: 1008335
Conc: 32.21 ng/ml




#3 n-Heptane

R.T.: 7.758 min
Delta R.T.: 0.002 min
Response: 333426
Conc: 10.72 ng/ml



#4 Benzene


R.T.: 7.893 min
Delta R.T.: 0.000 min
Response: 405263
Conc: 10.63 ng/ml



#5 AAA-TFT


R.T.: 8.799 min
Delta R.T.: 0.001 min
Response: 492858
Conc: 21.40 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.631 min
Delta R.T.: 0.002 min
Response: 1138780
Conc: 31.14 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.001 min
Response: 345122
Conc: 10.10 ng/ml



#8 m-Xylene


R.T.: 13.203 min
Delta R.T.: 0.001 min
Response: 675996
Conc: 20.03 ng/ml



#9 O-Xylene

R.T.: 13.931 min
Delta R.T.: 0.002 min
Response: 664449
Conc: 20.22 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.208 min
Delta R.T.: 0.000 min
Response: 539763
Conc: 20.01 ng/ml