

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080818\
 Data File : FB017208.D
 Signal(s) : FID2B.CH
 Acq On : 8 Aug 2018 12:05
 Operator : MJ/SJ
 Sample : FB080818GROICV
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 08 14:20:15 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
 Quant Title :
 QLast Update : Wed Aug 08 11:39:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.118	183123	19.545 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.968	495028	28.950 ng/ml
2) t 2,2,4-Trimethylpentane	7.734	493140	29.187 ng/ml
3) t n-Heptane	8.072	162232	10.071 ng/ml
4) t Benzene	8.212	202789	10.007 ng/ml
6) t Toluene	10.967	492768	29.895 ng/ml
7) t Ethylbenzene	13.416	158092	9.476 ng/ml
8) t m-Xylene	13.549	332343	17.728 ng/ml
9) t O-Xylene	14.280	348891	17.167 ng/ml
10) t 1,2,4-Trimethylbenzene	16.563	430194	19.209 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080818\
Data File : FB017208.D
Signal(s) : FID2B.CH
Acq On : 8 Aug 2018 12:05
Operator : MJ/SJ
Sample : FB080818GROICV
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 08 14:20:15 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
Quant Title :
QLast Update : Wed Aug 08 11:39:03 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.968 min
Delta R.T.: 0.004 min
Response: 495028
Conc: 28.95 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.734 min
Delta R.T.: 0.005 min
Response: 493140
Conc: 29.19 ng/ml




#3 n-Heptane

R.T.: 8.072 min
Delta R.T.: 0.004 min
Response: 162232
Conc: 10.07 ng/ml




#4 Benzene

R.T.: 8.212 min
Delta R.T.: 0.004 min
Response: 202789
Conc: 10.01 ng/ml




#5 AAA-TFT

R.T.: 9.118 min
Delta R.T.: 0.005 min
Response: 183123
Conc: 19.54 ng/ml




#6 Toluene

R.T.: 10.967 min
Delta R.T.: 0.005 min
Response: 492768
Conc: 29.89 ng/ml




#7 Ethylbenzene

R.T.: 13.416 min
Delta R.T.: 0.004 min
Response: 158092
Conc: 9.48 ng/ml




#8 m-Xylene

R.T.: 13.549 min
Delta R.T.: 0.004 min
Response: 332343
Conc: 17.73 ng/ml



#9 O-Xylene

R.T.: 14.280 min
Delta R.T.: 0.004 min
Response: 348891
Conc: 17.17 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.563 min
Delta R.T.: 0.004 min
Response: 430194
Conc: 19.21 ng/ml