

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080818\
Data File : FB017209.D
Signal(s) : FID2B.CH
Acq On : 8 Aug 2018 12:35
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 09 10:31:17 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
Quant Title :
QLast Update : Wed Aug 08 11:39:03 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.116	195255	20.840 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.969	530110	31.001 ng/ml
2) t 2,2,4-Trimethylpentane	7.732	519304	30.735 ng/ml
3) t n-Heptane	8.072	164876	10.235 ng/ml
4) t Benzene	8.211	208766	10.302 ng/ml
6) t Toluene	10.966	510406	30.965 ng/ml
7) t Ethylbenzene	13.416	163280	9.787 ng/ml
8) t m-Xylene	13.548	344561	18.380 ng/ml
9) t O-Xylene	14.279	352428	17.341 ng/ml
10) t 1,2,4-Trimethylbenzene	16.561	438205	19.566 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080818\
Data File : FB017209.D
Signal(s) : FID2B.CH
Acq On : 8 Aug 2018 12:35
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 09 10:31:17 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
Quant Title :
QLast Update : Wed Aug 08 11:39:03 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.969 min
Delta R.T.: 0.005 min
Response: 530110
Conc: 31.00 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.732 min
Delta R.T.: 0.003 min
Response: 519304
Conc: 30.74 ng/ml




#3 n-Heptane

R.T.: 8.072 min
Delta R.T.: 0.004 min
Response: 164876
Conc: 10.23 ng/ml




#4 Benzene

R.T.: 8.211 min
Delta R.T.: 0.004 min
Response: 208766
Conc: 10.30 ng/ml




#5 AAA-TFT

R.T.: 9.116 min
Delta R.T.: 0.004 min
Response: 195255
Conc: 20.84 ng/ml




#6 Toluene

R.T.: 10.966 min
Delta R.T.: 0.004 min
Response: 510406
Conc: 30.96 ng/ml




#7 Ethylbenzene

R.T.: 13.416 min
Delta R.T.: 0.004 min
Response: 163280
Conc: 9.79 ng/ml




#8 m-Xylene

R.T.: 13.548 min
Delta R.T.: 0.003 min
Response: 344561
Conc: 18.38 ng/ml



#9 O-Xylene

R.T.: 14.279 min
Delta R.T.: 0.003 min
Response: 352428
Conc: 17.34 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.561 min
Delta R.T.: 0.002 min
Response: 438205
Conc: 19.57 ng/ml