

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080918\
 Data File : FB017242.D
 Signal(s) : FID2B.CH
 Acq On : 9 Aug 2018 20:21
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 10 09:28:24 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
 Quant Title :
 QLast Update : Wed Aug 08 11:39:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.116	232502	24.815 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.967	567920	33.212 ng/ml
2) t 2,2,4-Trimethylpentane	7.733	554515	32.819 ng/ml
3) t n-Heptane	8.073	176051	10.928 ng/ml
4) t Benzene	8.211	217867	10.751 ng/ml
6) t Toluene	10.966	534120	32.403 ng/ml
7) t Ethylbenzene	13.416	169477	10.158 ng/ml
8) t m-Xylene	13.549	356935	19.040 ng/ml
9) t O-Xylene	14.281	373078	18.357 ng/ml
10) t 1,2,4-Trimethylbenzene	16.563	456526	20.384 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB080918\
Data File : FB017242.D
Signal(s) : FID2B.CH
Acq On : 9 Aug 2018 20:21
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 23 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 10 09:28:24 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
Quant Title :
QLast Update : Wed Aug 08 11:39:03 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.967 min
Delta R.T.: 0.003 min
Response: 567920
Conc: 33.21 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.733 min
Delta R.T.: 0.005 min
Response: 554515
Conc: 32.82 ng/ml




#3 n-Heptane

R.T.: 8.073 min
Delta R.T.: 0.005 min
Response: 176051
Conc: 10.93 ng/ml




#4 Benzene

R.T.: 8.211 min
Delta R.T.: 0.004 min
Response: 217867
Conc: 10.75 ng/ml




#5 AAA-TFT

R.T.: 9.116 min
Delta R.T.: 0.004 min
Response: 232502
Conc: 24.81 ng/ml




#6 Toluene

R.T.: 10.966 min
Delta R.T.: 0.004 min
Response: 534120
Conc: 32.40 ng/ml




#7 Ethylbenzene

R.T.: 13.416 min
Delta R.T.: 0.004 min
Response: 169477
Conc: 10.16 ng/ml




#8 m-Xylene

R.T.: 13.549 min
Delta R.T.: 0.004 min
Response: 356935
Conc: 19.04 ng/ml



#9 O-Xylene

R.T.: 14.281 min
Delta R.T.: 0.004 min
Response: 373078
Conc: 18.36 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.563 min
Delta R.T.: 0.004 min
Response: 456526
Conc: 20.38 ng/ml