

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB080922\  
 Data File : FB027887.D  
 Signal(s) : FID2B.CH  
 Acq On : 9 Aug 2022 12:36  
 Operator : AJ/MA  
 Sample : FB080922GROICV  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 FID\_B  
 ClientSampleId :  
 FB080922GROICV

Integration File: Calibration.e  
 Quant Time: Aug 09 12:49:35 2022  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB080922.M  
 Quant Title :  
 QLast Update : Tue Aug 09 12:12:39 2022  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 8.801  | 235945   | 20.359 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.720  | 542628   | 29.995 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.438  | 595638   | 30.374 ng/ml |
| 3) t n-Heptane               | 7.759  | 190497   | 9.888 ng/ml  |
| 4) t Benzene                 | 7.896  | 225364   | 9.693 ng/ml  |
| 6) t Toluene                 | 10.628 | 665051   | 30.329 ng/ml |
| 7) t Ethylbenzene            | 13.065 | 209407   | 10.162 ng/ml |
| 8) t m-Xylene                | 13.198 | 415555   | 20.308 ng/ml |
| 9) t O-Xylene                | 13.925 | 402671   | 20.328 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.200 | 396728   | 20.344 ng/ml |
| -----                        |        |          |              |

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB080922\  
Data File : FB027887.D  
Signal(s) : FID2B.CH  
Acq On : 9 Aug 2022 12:36  
Operator : AJ/MA  
Sample : FB080922GROICV  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
FID\_B  
ClientSampleId :  
FB080922GROICV

Integration File: Calibration.e  
Quant Time: Aug 09 12:49:35 2022  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB080922.M  
Quant Title :  
QLast Update : Tue Aug 09 12:12:39 2022  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.720 min  
Delta R.T.: 0.005 min  
Response: 542628  
Conc: 29.99 ng/ml

Instrument :  
FID\_B  
ClientSampleId :  
FB080922GROICV




#2 2,2,4-Trimethylpentane

R.T.: 7.438 min  
Delta R.T.: -0.005 min  
Response: 595638  
Conc: 30.37 ng/ml




#3 n-Heptane

R.T.: 7.759 min  
Delta R.T.: 0.004 min  
Response: 190497  
Conc: 9.89 ng/ml



#4 Benzene


R.T.: 7.896 min  
Delta R.T.: 0.004 min  
Response: 225364  
Conc: 9.69 ng/ml



#5 AAA-TFT


R.T.: 8.801 min  
Delta R.T.: 0.003 min  
Response: 235945  
Conc: 20.36 ng/ml

Instrument :  
FID\_B  
ClientSampleId :  
FB080922GROICV




#6 Toluene

R.T.: 10.628 min  
Delta R.T.: 0.001 min  
Response: 665051  
Conc: 30.33 ng/ml




#7 Ethylbenzene

R.T.: 13.065 min  
Delta R.T.: 0.000 min  
Response: 209407  
Conc: 10.16 ng/ml



#8 m-Xylene


R.T.: 13.198 min  
Delta R.T.: 0.001 min  
Response: 415555  
Conc: 20.31 ng/ml



#9 O-Xylene

R.T.: 13.925 min  
Delta R.T.: 0.000 min  
Response: 402671  
Conc: 20.33 ng/ml

Instrument :  
FID\_B  
ClientSampleId :  
FB080922GROICV



#10 1,2,4-Trimethylbenzene

R.T.: 16.200 min  
Delta R.T.: 0.000 min  
Response: 396728  
Conc: 20.34 ng/ml