

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081419\
 Data File : FB022009.D
 Signal(s) : FID2B.CH
 Acq On : 14 Aug 2019 10:35
 Operator : DD/AJ
 Sample : BSF0814S1
 Misc : 5.02/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 15 08:07:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB071519.M
 Quant Title :
 QLast Update : Mon Jul 15 13:30:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.989	272140	19.942 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.862	813406	29.497 ng/ml
2) t 2,2,4-Trimethylpentane	7.611	831632	31.460 ng/ml
3) t n-Heptane	7.950	257256	10.376 ng/ml
4) t Benzene	8.084	294564	10.083 ng/ml
6) t Toluene	10.831	625884	30.418 ng/ml
7) t Ethylbenzene	13.277	212421	10.210 ng/ml
8) t m-Xylene	13.409	447445	20.141 ng/ml
9) t O-Xylene	14.139	468560	20.370 ng/ml
10) t 1,2,4-Trimethylbenzene	16.418	484076	20.532 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081419\
Data File : FB022009.D
Signal(s) : FID2B.CH
Acq On : 14 Aug 2019 10:35
Operator : DD/AJ
Sample : BSF0814S1
Misc : 5.02/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 15 08:07:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB071519.M
Quant Title :
QLast Update : Mon Jul 15 13:30:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

