

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB081423\  
Data File : FB030264.D  
Signal(s) : FID2B.CH  
Acq On : 15 Aug 2023 00:55  
Operator : YP/AJ  
Sample : BSF0814S3  
Misc : 5.00G/5.00 ML DI WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
FID\_B  
ClientSampleId :  
BSF0814S3

Integration File: Calibration.e  
Quant Time: Aug 15 02:31:03 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB072423.M  
Quant Title :  
QLast Update : Tue Jul 25 06:41:02 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um

| Compound                    |                        | R.T.   | Response | Conc Units   |
|-----------------------------|------------------------|--------|----------|--------------|
| -----                       |                        |        |          |              |
| System Monitoring Compounds |                        |        |          |              |
| 5) s                        | AAA-TFT                | 8.792  | 193886   | 15.775 ng/ml |
| Target Compounds            |                        |        |          |              |
| 1) t                        | 2-Methylpentane        | 4.708  | 412162   | 24.950 ng/ml |
| 2) t                        | 2,2,4-Trimethylpentane | 7.419  | 520020   | 28.024 ng/ml |
| 3) t                        | n-Heptane              | 7.751  | 157300   | 8.176 ng/ml  |
| 4) t                        | Benzene                | 7.888  | 216787   | 9.215 ng/ml  |
| 6) t                        | Toluene                | 10.620 | 644183   | 28.109 ng/ml |
| 7) t                        | Ethylbenzene           | 13.057 | 207570   | 9.671 ng/ml  |
| 8) t                        | m-Xylene               | 13.189 | 404096   | 19.215 ng/ml |
| 9) t                        | O-Xylene               | 13.916 | 389502   | 18.978 ng/ml |
| 10) t                       | 1,2,4-Trimethylbenzene | 16.191 | 371017   | 18.955 ng/ml |
| -----                       |                        |        |          |              |

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB081423\  
Data File : FB030264.D  
Signal(s) : FID2B.CH  
Acq On : 15 Aug 2023 00:55  
Operator : YP/AJ  
Sample : BSF0814S3  
Misc : 5.00G/5.00 ML DI WATER  
ALS Vial : 25 Sample Multiplier: 1

Instrument :  
FID\_B  
ClientSampleId :  
BSF0814S3

Integration File: Calibration.e  
Quant Time: Aug 15 02:31:03 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB072423.M  
Quant Title :  
QLast Update : Tue Jul 25 06:41:02 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um

