

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081619\
 Data File : FB022044.D
 Signal(s) : FID2B.CH
 Acq On : 16 Aug 2019 20:01
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 17 00:33:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB081519.M
 Quant Title :
 QLast Update : Thu Aug 15 11:19:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.990	276445	21.411 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.862	820838	30.294 ng/ml
2) t 2,2,4-Trimethylpentane	7.611	821167	30.776 ng/ml
3) t n-Heptane	7.950	257581	10.546 ng/ml
4) t Benzene	8.085	295897	10.262 ng/ml
6) t Toluene	10.831	620515	31.263 ng/ml
7) t Ethylbenzene	13.276	207534	10.291 ng/ml
8) t m-Xylene	13.409	438928	20.274 ng/ml
9) t O-Xylene	14.139	456725	20.394 ng/ml
10) t 1,2,4-Trimethylbenzene	16.418	474996	20.218 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081619\
Data File : FB022044.D
Signal(s) : FID2B.CH
Acq On : 16 Aug 2019 20:01
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 17 00:33:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB081519.M
Quant Title :
QLast Update : Thu Aug 15 11:19:59 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.862 min
Delta R.T.: 0.000 min
Response: 820838
Conc: 30.29 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.611 min
Delta R.T.: 0.000 min
Response: 821167
Conc: 30.78 ng/ml




#3 n-Heptane

R.T.: 7.950 min
Delta R.T.: -0.001 min
Response: 257581
Conc: 10.55 ng/ml




#4 Benzene

R.T.: 8.085 min
Delta R.T.: -0.001 min
Response: 295897
Conc: 10.26 ng/ml




#5 AAA-TFT

R.T.: 8.990 min
Delta R.T.: -0.002 min
Response: 276445
Conc: 21.41 ng/ml




#6 Toluene

R.T.: 10.831 min
Delta R.T.: 0.000 min
Response: 620515
Conc: 31.26 ng/ml




#7 Ethylbenzene

R.T.: 13.276 min
Delta R.T.: -0.002 min
Response: 207534
Conc: 10.29 ng/ml



#8 m-Xylene

R.T.: 13.409 min
Delta R.T.: -0.001 min
Response: 438928
Conc: 20.27 ng/ml




#9 O-Xylene

R.T.: 14.139 min

Delta R.T.: -0.001 min

Response: 456725

Conc: 20.39 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.418 min

Delta R.T.: -0.001 min

Response: 474996

Conc: 20.22 ng/ml