

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081922\
 Data File : FB028004.D
 Signal(s) : FID2B.CH
 Acq On : 19 Aug 2022 13:04
 Operator : AJ/MA
 Sample : BSF0819S1
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0819S1

Integration File: Calibration.e
 Quant Time: Aug 19 16:08:32 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080922.M
 Quant Title :
 QLast Update : Tue Aug 09 12:12:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	248254	21.421 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.716	590606	32.647 ng/ml
2) t 2,2,4-Trimethylpentane	7.432	612405	31.229 ng/ml
3) t n-Heptane	7.755	206647	10.726 ng/ml
4) t Benzene	7.892	237095	10.197 ng/ml
6) t Toluene	10.624	672219	30.656 ng/ml
7) t Ethylbenzene	13.061	210807	10.230 ng/ml
8) t m-Xylene	13.194	418193	20.437 ng/ml
9) t O-Xylene	13.921	403090	20.349 ng/ml
10) t 1,2,4-Trimethylbenzene	16.196	388719	19.933 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081922\
Data File : FB028004.D
Signal(s) : FID2B.CH
Acq On : 19 Aug 2022 13:04
Operator : AJ/MA
Sample : BSF0819S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0819S1

Integration File: Calibration.e
Quant Time: Aug 19 16:08:32 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080922.M
Quant Title :
QLast Update : Tue Aug 09 12:12:39 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

