

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082124\
 Data File : FB030882.D
 Signal(s) : FID2B.CH
 Acq On : 21 Aug 2024 12:22
 Operator : YP/AJ
 Sample : BSF0821S2
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF0821S2

Integration File: Calibration.e
 Quant Time: Aug 22 02:57:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB073024.M
 Quant Title :
 QLast Update : Tue Jul 30 12:33:32 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.789	200018	16.987 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.708	464541	22.052 ng/ml
2) t 2,2,4-Trimethylpentane	7.412	540982	25.395 ng/ml
3) t n-Heptane	7.747	146606	6.601 ng/ml
4) t Benzene	7.883	189949	7.203 ng/ml
6) t Toluene	10.616	559249	22.011 ng/ml
7) t Ethylbenzene	13.052	186422	7.886 ng/ml
8) t m-Xylene	13.185	380128	16.102 ng/ml
9) t O-Xylene	13.912	379837	16.491 ng/ml
10) t 1,2,4-Trimethylbenzene	16.186	398883	17.404 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082124\
Data File : FB030882.D
Signal(s) : FID2B.CH
Acq On : 21 Aug 2024 12:22
Operator : YP/AJ
Sample : BSF0821S2
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF0821S2

Integration File: Calibration.e
Quant Time: Aug 22 02:57:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB073024.M
Quant Title :
QLast Update : Tue Jul 30 12:33:32 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

