

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB083023\
Data File : FB030298.D
Signal(s) : FID2B.CH
Acq On : 30 Aug 2023 9:25
Operator : YP/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 30 09:29:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082223.M
Quant Title :
QLast Update : Tue Aug 22 18:06:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 t	2-Methylpentane	18.257	19.847 E3	-8.7	110	0.00
2 t	2,2,4-Trimethylpentane	19.731	21.670 E3	-9.8	113	0.00
3 t	n-Heptane	21.319	23.116 E3	-8.4	108	0.00
4 t	Benzene	24.840	26.892 E3	-8.3	107	0.00
5 s	AAA-TFT	11.901	13.723 E3	-15.3	103	0.00
6 t	Toluene	23.980	26.272 E3	-9.6	108	0.00
7 t	Ethylbenzene	22.418	24.319 E3	-8.5	108	0.00
8 t	m-Xylene	22.296	24.091 E3	-8.1	108	0.00
9 t	O-Xylene	21.328	23.318 E3	-9.3	110	0.00
10 t	1,2,4-Trimethylbenzene	21.190	22.482 E3	-6.1	111	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0