

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082419\
 Data File : FB022173.D
 Signal(s) : FID2B.CH
 Acq On : 24 Aug 2019 20:02
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: Calibration.e
 Quant Time: Aug 26 08:10:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
 Quant Title :
 QLast Update : Wed Aug 21 07:59:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.983	437251	20.586 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.854	1220270	30.490 ng/ml
2) t 2,2,4-Trimethylpentane	7.603	1216004	30.190 ng/ml
3) t n-Heptane	7.942	381035	9.975 ng/ml
4) t Benzene	8.077	459692	10.240 ng/ml
6) t Toluene	10.824	937139	30.484 ng/ml
7) t Ethylbenzene	13.270	208707	10.227 ng/ml
8) t m-Xylene	13.403	430640	20.448 ng/ml
9) t O-Xylene	14.132	457296	20.447 ng/ml
10) t 1,2,4-Trimethylbenzene	16.411	511305	19.937 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082419\
Data File : FB022173.D
Signal(s) : FID2B.CH
Acq On : 24 Aug 2019 20:02
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: Calibration.e
Quant Time: Aug 26 08:10:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
Quant Title :
QLast Update : Wed Aug 21 07:59:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.854 min
Delta R.T.: 0.002 min
Response: 1220270
Conc: 30.49 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.603 min
Delta R.T.: 0.003 min
Response: 1216004
Conc: 30.19 ng/ml




#3 n-Heptane

R.T.: 7.942 min
Delta R.T.: 0.001 min
Response: 381035
Conc: 9.97 ng/ml



#4 Benzene


R.T.: 8.077 min
Delta R.T.: 0.000 min
Response: 459692
Conc: 10.24 ng/ml



#5 AAA-TFT


R.T.: 8.983 min
Delta R.T.: 0.001 min
Response: 437251
Conc: 20.59 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#6 Toluene

R.T.: 10.824 min
Delta R.T.: 0.001 min
Response: 937139
Conc: 30.48 ng/ml




#7 Ethylbenzene

R.T.: 13.270 min
Delta R.T.: 0.000 min
Response: 208707
Conc: 10.23 ng/ml



#8 m-Xylene


R.T.: 13.403 min
Delta R.T.: 0.000 min
Response: 430640
Conc: 20.45 ng/ml



#9 O-Xylene

R.T.: 14.132 min
Delta R.T.: 0.000 min
Response: 457296
Conc: 20.45 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.411 min
Delta R.T.: 0.000 min
Response: 511305
Conc: 19.94 ng/ml