

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082420\
 Data File : FB024793.D
 Signal(s) : FID2B.CH
 Acq On : 24 Aug 2020 17:46
 Operator : DD/AJ
 Sample : L3773-03DL 50X
 Misc : 5.01g/5.00ml DI WATER
 ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 25 00:54:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB081020.M
 Quant Title :
 QLast Update : Mon Aug 10 12:11:06 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.801	414870	20.091 ng/ml
Target Compounds			
2) t 2,2,4-Trimethylpentane	7.425	1973	0.071 ng/ml
4) t Benzene	7.924	1796	0.051 ng/ml
6) t Toluene	10.630	1413	0.041 ng/ml
7) t Ethylbenzene	13.038	56089	1.752 ng/ml
8) t m-Xylene	13.202	16205	0.519 ng/ml
9) t o-Xylene	13.933	44956	1.487 ng/ml
10) t 1,2,4-Trimethylbenzene	16.212	329666	13.423 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082420\
Data File : FB024793.D
Signal(s) : FID2B.CH
Acq On : 24 Aug 2020 17:46
Operator : DD/AJ
Sample : L3773-03DL 50X
Misc : 5.01g/5.00ml DI WATER
ALS Vial : 12 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 25 00:54:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB081020.M
Quant Title :
QLast Update : Mon Aug 10 12:11:06 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#2 2,2,4-Trimethylpentane

R.T.: 7.425 min
Delta R.T.: 0.008 min
Response: 1973
Conc: 0.07 ng/ml




#4 Benzene

R.T.: 7.924 min
Delta R.T.: 0.032 min
Response: 1796
Conc: 0.05 ng/ml




#5 AAA-TFT

R.T.: 8.801 min
Delta R.T.: 0.002 min
Response: 414870
Conc: 20.09 ng/ml




#6 Toluene

R.T.: 10.630 min
Delta R.T.: 0.000 min
Response: 1413
Conc: 0.04 ng/ml




#7 Ethylbenzene

R.T.: 13.038 min
Delta R.T.: -0.031 min
Response: 56089
Conc: 1.75 ng/ml




#8 m-Xylene

R.T.: 13.202 min
Delta R.T.: 0.000 min
Response: 16205
Conc: 0.52 ng/ml



#9 O-Xylene

R.T.: 13.933 min
Delta R.T.: 0.004 min
Response: 44956
Conc: 1.49 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.212 min
Delta R.T.: 0.006 min
Response: 329666
Conc: 13.42 ng/ml