

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082719\
 Data File : FB022205.D
 Signal(s) : FID2B.CH
 Acq On : 27 Aug 2019 12:32
 Operator : DD/AJ
 Sample : K4093-04MS
 Misc : 5.04/5.00ml DI WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 PL-02-082619-1MS

Integration File: Calibration.e
 Quant Time: Aug 27 23:17:04 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
 Quant Title :
 QLast Update : Wed Aug 21 07:59:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.985	438455	20.643 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.854	1035338	25.869 ng/ml
2) t 2,2,4-Trimethylpentane	7.605	897295	22.278 ng/ml
3) t n-Heptane	7.944	276499	7.238 ng/ml
4) t Benzene	8.079	426227	9.494 ng/ml
6) t Toluene	10.826	839180	27.297 ng/ml
7) t Ethylbenzene	13.272	195295	9.570 ng/ml
8) t m-Xylene	13.405	400728	19.028 ng/ml
9) t O-Xylene	14.134	415253	18.567 ng/ml
10) t 1,2,4-Trimethylbenzene	16.414	439165	17.124 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082719\
Data File : FB022205.D
Signal(s) : FID2B.CH
Acq On : 27 Aug 2019 12:32
Operator : DD/AJ
Sample : K4093-04MS
Misc : 5.04/5.00ml DI WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
PL-02-082619-1MS

Integration File: Calibration.e
Quant Time: Aug 27 23:17:04 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
Quant Title :
QLast Update : Wed Aug 21 07:59:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

