

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB083023\
 Data File : FB030309.D
 Signal(s) : FID2B.CH
 Acq On : 30 Aug 2023 16:40
 Operator : YP/AJ
 Sample : 04217-01MSD 50X
 Misc : 5.10G/5.00 ML MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Aug 31 00:42:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082223.M
 Quant Title :
 QLast Update : Tue Aug 22 18:06:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	272687	22.912 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.698	487769	26.819 ng/ml
2) t 2,2,4-Trimethylpentane	7.421	556297	28.194 ng/ml
3) t n-Heptane	7.748	196732	9.228 ng/ml
4) t Benzene	7.884	240442	9.679 ng/ml
6) t Toluene	10.622	721963	30.107 ng/ml
7) t Ethylbenzene	13.060	220091	9.818 ng/ml
8) t m-Xylene	13.193	434682	19.496 ng/ml
9) t O-Xylene	13.920	416080	19.508 ng/ml
10) t 1,2,4-Trimethylbenzene	16.195	367836	17.359 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB083023\
Data File : FB030309.D
Signal(s) : FID2B.CH
Acq On : 30 Aug 2023 16:40
Operator : YP/AJ
Sample : 04217-01MSD 50X
Misc : 5.10G/5.00 ML MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Aug 31 00:42:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082223.M
Quant Title :
QLast Update : Tue Aug 22 18:06:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.698 min
Delta R.T.: -0.008 min
Response: 487769
Conc: 26.82 ng/ml

Instrument :


FID_B

ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.421 min
Delta R.T.: 0.005 min
Response: 556297
Conc: 28.19 ng/ml




#3 n-Heptane

R.T.: 7.748 min
Delta R.T.: 0.000 min
Response: 196732
Conc: 9.23 ng/ml



#4 Benzene


R.T.: 7.884 min
Delta R.T.: 0.000 min
Response: 240442
Conc: 9.68 ng/ml



#5 AAA-TFT


R.T.: 8.791 min
Delta R.T.: 0.002 min
Response: 272687
Conc: 22.91 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.622 min
Delta R.T.: 0.005 min
Response: 721963
Conc: 30.11 ng/ml




#7 Ethylbenzene

R.T.: 13.060 min
Delta R.T.: 0.008 min
Response: 220091
Conc: 9.82 ng/ml



#8 m-Xylene


R.T.: 13.193 min
Delta R.T.: 0.009 min
Response: 434682
Conc: 19.50 ng/ml



#9 O-Xylene

R.T.: 13.920 min
Delta R.T.: 0.009 min
Response: 416080
Conc: 19.51 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.195 min
Delta R.T.: 0.009 min
Response: 367836
Conc: 17.36 ng/ml