

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090419\
 Data File : FB022272.D
 Signal(s) : FID2B.CH
 Acq On : 4 Sep 2019 14:37
 Operator : DD/AJ
 Sample : 5 GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Sep 04 16:01:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
 Quant Title :
 QLast Update : Wed Sep 04 16:00:37 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.980	129971	4.457 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.852	336680	6.982 ng/ml
2) t 2,2,4-Trimethylpentane	7.600	338320	6.553 ng/ml
3) t n-Heptane	7.939	110204	2.183 ng/ml
4) t Benzene	8.073	127589	2.097 ng/ml
6) t Toluene	10.821	291921	5.734 ng/ml
7) t Ethylbenzene	13.266	62332	1.872 ng/ml
8) t m-Xylene	13.399	116161	3.990 ng/ml
9) t O-Xylene	14.130	119671	4.739 ng/ml
10) t 1,2,4-Trimethylbenzene	16.409	133721	5.234 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090419\
Data File : FB022272.D
Signal(s) : FID2B.CH
Acq On : 4 Sep 2019 14:37
Operator : DD/AJ
Sample : 5 GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Sep 04 16:01:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
Quant Title :
QLast Update : Wed Sep 04 16:00:37 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.852 min
Delta R.T.: 0.005 min
Response: 336680
Conc: 6.98 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.600 min
Delta R.T.: 0.002 min
Response: 338320
Conc: 6.55 ng/ml




#3 n-Heptane

R.T.: 7.939 min
Delta R.T.: 0.000 min
Response: 110204
Conc: 2.18 ng/ml



#4 Benzene


R.T.: 8.073 min
Delta R.T.: 0.000 min
Response: 127589
Conc: 2.10 ng/ml



#5 AAA-TFT


R.T.: 8.980 min
Delta R.T.: 0.000 min
Response: 129971
Conc: 4.46 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.821 min
Delta R.T.: 0.002 min
Response: 291921
Conc: 5.73 ng/ml




#7 Ethylbenzene

R.T.: 13.266 min
Delta R.T.: 0.001 min
Response: 62332
Conc: 1.87 ng/ml



#8 m-Xylene


R.T.: 13.399 min
Delta R.T.: 0.001 min
Response: 116161
Conc: 3.99 ng/ml



#9 O-Xylene

R.T.: 14.130 min
Delta R.T.: 0.002 min
Response: 119671
Conc: 4.74 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.409 min
Delta R.T.: 0.002 min
Response: 133721
Conc: 5.23 ng/ml