

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090623\
 Data File : FB030329.D
 Signal(s) : FID2B.CH
 Acq On : 6 Sep 2023 22:55
 Operator : YP/AJ
 Sample : 04299-08MSD 50X
 Misc : 5.09G/5.00 ML MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 SB-04-COMPMSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/07/2023
 Supervised By :Ankita Jodhani 09/07/2023

Integration File: Calibration.e
 Quant Time: Sep 07 01:13:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082223.M
 Quant Title :
 QLast Update : Tue Aug 22 18:06:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.796	219246	18.422 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.709	424293	23.329 ng/ml
2) t 2,2,4-Trimethylpentane	7.426	480229	24.339 ng/ml
3) t n-Heptane	7.753	163701	7.679 ng/ml
4) t Benzene	7.891	640117	25.769 ng/ml
6) t Toluene	10.625	692172	28.864 ng/ml
7) t Ethylbenzene	13.062	1547537	69.031 ng/ml
8) t m-Xylene	13.192	1673817	75.073 ng/ml
9) t O-Xylene	13.920	1186734	55.642 ng/mlm
10) t 1,2,4-Trimethylbenzene	16.196	1199033	56.584 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090623\
Data File : FB030329.D
Signal(s) : FID2B.CH
Acq On : 6 Sep 2023 22:55
Operator : YP/AJ
Sample : 04299-08MSD 50X
Misc : 5.09G/5.00 ML MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
SB-04-COMPMSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/07/2023
Supervised By :Ankita Jodhani 09/07/2023

Integration File: Calibration.e
Quant Time: Sep 07 01:13:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082223.M
Quant Title :
QLast Update : Tue Aug 22 18:06:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

