

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090721\
 Data File : FB026216.D
 Signal(s) : FID2B.CH
 Acq On : 7 Sep 2021 13:29
 Operator : AJ/MA
 Sample : M3606-03MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Sep 08 02:03:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB081321.M
 Quant Title :
 QLast Update : Fri Aug 13 22:11:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.847	344060	23.089 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.855	634318	28.097 ng/ml
2) t 2,2,4-Trimethylpentane	7.509	669512	30.455 ng/ml
3) t n-Heptane	7.825	263839	8.758 ng/ml
4) t Benzene	7.963	295453	9.351 ng/ml
6) t Toluene	10.656	832411	28.917 ng/ml
7) t Ethylbenzene	13.074	233274	9.663 ng/ml
8) t m-Xylene	13.206	450419	19.272 ng/ml
9) t O-Xylene	13.930	407755	19.599 ng/ml
10) t 1,2,4-Trimethylbenzene	16.198	419359	20.788 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090721\
Data File : FB026216.D
Signal(s) : FID2B.CH
Acq On : 7 Sep 2021 13:29
Operator : AJ/MA
Sample : M3606-03MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Sep 08 02:03:55 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB081321.M
Quant Title :
QLast Update : Fri Aug 13 22:11:30 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.855 min
Delta R.T.: 0.013 min
Response: 634318
Conc: 28.10 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.509 min
Delta R.T.: -0.006 min
Response: 669512
Conc: 30.45 ng/ml




#3 n-Heptane

R.T.: 7.825 min
Delta R.T.: 0.003 min
Response: 263839
Conc: 8.76 ng/ml



#4 Benzene


R.T.: 7.963 min
Delta R.T.: 0.003 min
Response: 295453
Conc: 9.35 ng/ml



#5 AAA-TFT


R.T.: 8.847 min
Delta R.T.: 0.000 min
Response: 344060
Conc: 23.09 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.656 min
Delta R.T.: -0.002 min
Response: 832411
Conc: 28.92 ng/ml




#7 Ethylbenzene

R.T.: 13.074 min
Delta R.T.: -0.003 min
Response: 233274
Conc: 9.66 ng/ml



#8 m-Xylene


R.T.: 13.206 min
Delta R.T.: -0.003 min
Response: 450419
Conc: 19.27 ng/ml



#9 O-Xylene

R.T.: 13.930 min
Delta R.T.: -0.003 min
Response: 407755
Conc: 19.60 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.198 min
Delta R.T.: -0.004 min
Response: 419359
Conc: 20.79 ng/ml