

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090920\
 Data File : FB024845.D
 Signal(s) : FID2B.CH
 Acq On : 9 Sep 2020 12:30
 Operator : DD/AJ
 Sample : 10 GRO STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Sep 09 14:17:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090920.M
 Quant Title :
 QLast Update : Wed Sep 09 14:16:34 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	242398	11.189 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.687	520836	17.313 ng/ml
2) t 2,2,4-Trimethylpentane	7.415	541144	17.475 ng/ml
3) t n-Heptane	7.755	175025	5.737 ng/ml
4) t Benzene	7.891	208158	5.293 ng/ml
6) t Toluene	10.628	604992	16.394 ng/ml
7) t Ethylbenzene	13.068	185619	5.518 ng/ml
8) t m-Xylene	13.201	371546	11.124 ng/ml
9) t O-Xylene	13.929	361441	10.985 ng/ml
10) t 1,2,4-Trimethylbenzene	16.206	296151	10.738 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090920\
Data File : FB024845.D
Signal(s) : FID2B.CH
Acq On : 9 Sep 2020 12:30
Operator : DD/AJ
Sample : 10 GRO STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 09 14:17:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090920.M
Quant Title :
QLast Update : Wed Sep 09 14:16:34 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.687 min
Delta R.T.: 0.000 min
Response: 520836
Conc: 17.31 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.415 min
Delta R.T.: -0.003 min
Response: 541144
Conc: 17.48 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: 0.000 min
Response: 175025
Conc: 5.74 ng/ml




#4 Benzene

R.T.: 7.891 min
Delta R.T.: 0.000 min
Response: 208158
Conc: 5.29 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: 0.000 min
Response: 242398
Conc: 11.19 ng/ml




#6 Toluene

R.T.: 10.628 min
Delta R.T.: 0.000 min
Response: 604992
Conc: 16.39 ng/ml




#7 Ethylbenzene

R.T.: 13.068 min
Delta R.T.: 0.000 min
Response: 185619
Conc: 5.52 ng/ml




#8 m-Xylene

R.T.: 13.201 min
Delta R.T.: 0.000 min
Response: 371546
Conc: 11.12 ng/ml



#9 O-Xylene

R.T.: 13.929 min
Delta R.T.: 0.000 min
Response: 361441
Conc: 10.99 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.206 min
Delta R.T.: 0.000 min
Response: 296151
Conc: 10.74 ng/ml