

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091124\
 Data File : FB030922.D
 Signal(s) : FID2B.CH
 Acq On : 11 Sep 2024 12:16
 Operator : YP/AJ
 Sample : P3245-18RE
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Sep 12 01:33:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082724.M
 Quant Title :
 QLast Update : Tue Aug 27 11:54:45 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.802	479900	39.664 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.701	254586	14.247 ng/ml
2) t 2,2,4-Trimethylpentane	7.398	712342	38.852 ng/ml
3) t n-Heptane	7.745	478830	25.891 ng/ml
4) t Benzene	7.879	191242	8.465 ng/ml
6) t Toluene	10.614	1866147	85.443 ng/ml
7) t Ethylbenzene	13.049	524265	25.044 ng/ml
8) t m-Xylene	13.180	1645303	79.296 ng/ml
9) t O-Xylene	13.909	628420	31.180 ng/ml
10) t 1,2,4-Trimethylbenzene	16.184	683393	34.249 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091124\
Data File : FB030922.D
Signal(s) : FID2B.CH
Acq On : 11 Sep 2024 12:16
Operator : YP/AJ
Sample : P3245-18RE
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Sep 12 01:33:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082724.M
Quant Title :
QLast Update : Tue Aug 27 11:54:45 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.701 min
Delta R.T.: -0.001 min
Response: 254586
Conc: 14.25 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.398 min
Delta R.T.: -0.013 min
Response: 712342
Conc: 38.85 ng/ml




#3 n-Heptane

R.T.: 7.745 min
Delta R.T.: -0.002 min
Response: 478830
Conc: 25.89 ng/ml



#4 Benzene


R.T.: 7.879 min
Delta R.T.: -0.002 min
Response: 191242
Conc: 8.46 ng/ml



#5 AAA-TFT


R.T.: 8.802 min
Delta R.T.: 0.016 min
Response: 479900
Conc: 39.66 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.614 min
Delta R.T.: 0.000 min
Response: 1866147
Conc: 85.44 ng/ml




#7 Ethylbenzene

R.T.: 13.049 min
Delta R.T.: 0.000 min
Response: 524265
Conc: 25.04 ng/ml



#8 m-Xylene


R.T.: 13.180 min
Delta R.T.: -0.002 min
Response: 1645303
Conc: 79.30 ng/ml



#9 O-Xylene

R.T.: 13.909 min
Delta R.T.: 0.000 min
Response: 628420
Conc: 31.18 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.184 min
Delta R.T.: 0.000 min
Response: 683393
Conc: 34.25 ng/ml