

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091219\
 Data File : FB022479.D
 Signal(s) : FID2B.CH
 Acq On : 12 Sep 2019 21:16
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Sep 13 01:33:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
 Quant Title :
 QLast Update : Wed Sep 04 17:01:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.967	542862	18.663 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.833	1059571	21.565 ng/ml
2) t 2,2,4-Trimethylpentane	7.585	1188527	23.017 ng/ml
3) t n-Heptane	7.925	390114	7.751 ng/ml
4) t Benzene	8.060	493523	8.316 ng/ml
6) t Toluene	10.808	1461696	30.725 ng/ml
7) t Ethylbenzene	13.255	465395	15.524 ng/ml
8) t m-Xylene	13.388	921494	34.687 ng/ml
9) t O-Xylene	14.117	873025	35.986 ng/ml
10) t 1,2,4-Trimethylbenzene	16.398	577766	22.161 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091219\
Data File : FB022479.D
Signal(s) : FID2B.CH
Acq On : 12 Sep 2019 21:16
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 27 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 13 01:33:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
Quant Title :
QLast Update : Wed Sep 04 17:01:40 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.833 min
Delta R.T.: -0.014 min
Response: 1059571
Conc: 21.57 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.585 min
Delta R.T.: -0.012 min
Response: 1188527
Conc: 23.02 ng/ml




#3 n-Heptane

R.T.: 7.925 min
Delta R.T.: -0.012 min
Response: 390114
Conc: 7.75 ng/ml




#4 Benzene

R.T.: 8.060 min
Delta R.T.: -0.012 min
Response: 493523
Conc: 8.32 ng/ml




#5 AAA-TFT

R.T.: 8.967 min
Delta R.T.: -0.012 min
Response: 542862
Conc: 18.66 ng/ml




#6 Toluene

R.T.: 10.808 min
Delta R.T.: -0.011 min
Response: 1461696
Conc: 30.73 ng/ml




#7 Ethylbenzene

R.T.: 13.255 min
Delta R.T.: -0.011 min
Response: 465395
Conc: 15.52 ng/ml




#8 m-Xylene

R.T.: 13.388 min
Delta R.T.: -0.010 min
Response: 921494
Conc: 34.69 ng/ml



#9 O-Xylene

R.T.: 14.117 min
Delta R.T.: -0.011 min
Response: 873025
Conc: 35.99 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.398 min
Delta R.T.: -0.009 min
Response: 577766
Conc: 22.16 ng/ml