

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091219\
Data File : FB022474.D
Signal(s) : FID2B.CH
Acq On : 12 Sep 2019 18:44
Operator : DD/AJ
Sample : K4857-20
Misc : 5.07g/5.00ML DI WATER
ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 13 01:31:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
Quant Title :
QLast Update : Wed Sep 04 17:01:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.968	483029	16.606 ng/ml
Target Compounds			


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091219\
Data File : FB022474.D
Signal(s) : FID2B.CH
Acq On : 12 Sep 2019 18:44
Operator : DD/AJ
Sample : K4857-20
Misc : 5.07g/5.00ML DI WATER
ALS Vial : 22 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 13 01:31:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
Quant Title :
QLast Update : Wed Sep 04 17:01:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um





#5 AAA-TFT

R.T.: 8.968 min

Delta R.T.: -0.011 min

Response: 483029

Conc: 16.61 ng/ml