

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091522\
Data File : FB028279.D
Signal(s) : FID2B.CH
Acq On : 15 Sep 2022 15:10
Operator : AJ/MA
Sample : N4630-19
Misc : 5.05G/5.00 ML DI WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Sep 16 05:53:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090822.M
Quant Title :
QLast Update : Thu Sep 08 15:28:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	88998	7.436 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.711	7575	0.374 ng/ml
7) t Ethylbenzene	13.037	29022	1.305 ng/ml
9) t O-Xylene	13.945	1973	0.094 ng/ml
10) t 1,2,4-Trimethylbenzene	16.222	5196	0.258 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091522\
Data File : FB028279.D
Signal(s) : FID2B.CH
Acq On : 15 Sep 2022 15:10
Operator : AJ/MA
Sample : N4630-19
Misc : 5.05G/5.00 ML DI WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Sep 16 05:53:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090822.M
Quant Title :
QLast Update : Thu Sep 08 15:28:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.711 min
Delta R.T.: -0.007 min
Response: 7575
Conc: 0.37 ng/ml

Instrument :
FID_B
ClientSampleId :




#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: -0.001 min
Response: 88998
Conc: 7.44 ng/ml




#7 Ethylbenzene

R.T.: 13.037 min
Delta R.T.: -0.027 min
Response: 29022
Conc: 1.31 ng/ml



#9 O-Xylene

R.T.: 13.945 min
Delta R.T.: 0.021 min
Response: 1973
Conc: 0.09 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.222 min
Delta R.T.: 0.023 min
Response: 5196
Conc: 0.26 ng/ml

Instrument :
FID_B
ClientSampleId :