

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091619\
 Data File : FB022535.D
 Signal(s) : FID2B.CH
 Acq On : 17 Sep 2019 00:33
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 20 PPB GRO STD

Integration File: Calibration.e
 Quant Time: Sep 17 02:20:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
 Quant Title :
 QLast Update : Wed Sep 04 17:01:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.967	542080	18.636 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.833	980555	19.957 ng/ml
2) t 2,2,4-Trimethylpentane	7.585	1075343	20.825 ng/ml
3) t n-Heptane	7.925	353641	7.026 ng/ml
4) t Benzene	8.059	460303	7.756 ng/ml
6) t Toluene	10.808	1360131	28.590 ng/ml
7) t Ethylbenzene	13.255	430815	14.370 ng/ml
8) t m-Xylene	13.388	852550	32.091 ng/ml
9) t O-Xylene	14.117	813057	33.514 ng/ml
10) t 1,2,4-Trimethylbenzene	16.398	551715	21.162 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB091619\
Data File : FB022535.D
Signal(s) : FID2B.CH
Acq On : 17 Sep 2019 00:33
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD

Integration File: Calibration.e
Quant Time: Sep 17 02:20:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
Quant Title :
QLast Update : Wed Sep 04 17:01:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.833 min
Delta R.T.: -0.014 min
Response: 980555
Conc: 19.96 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#2 2,2,4-Trimethylpentane

R.T.: 7.585 min
Delta R.T.: -0.013 min
Response: 1075343
Conc: 20.82 ng/ml




#3 n-Heptane

R.T.: 7.925 min
Delta R.T.: -0.013 min
Response: 353641
Conc: 7.03 ng/ml



#4 Benzene


R.T.: 8.059 min
Delta R.T.: -0.013 min
Response: 460303
Conc: 7.76 ng/ml



#5 AAA-TFT


R.T.: 8.967 min
Delta R.T.: -0.012 min
Response: 542080
Conc: 18.64 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD




#6 Toluene

R.T.: 10.808 min
Delta R.T.: -0.011 min
Response: 1360131
Conc: 28.59 ng/ml




#7 Ethylbenzene

R.T.: 13.255 min
Delta R.T.: -0.011 min
Response: 430815
Conc: 14.37 ng/ml



#8 m-Xylene


R.T.: 13.388 min
Delta R.T.: -0.010 min
Response: 852550
Conc: 32.09 ng/ml



#9 O-Xylene

R.T.: 14.117 min
Delta R.T.: -0.011 min
Response: 813057
Conc: 33.51 ng/ml

Instrument :
FID_B
ClientSampleId :
20 PPB GRO STD



#10 1,2,4-Trimethylbenzene

R.T.: 16.398 min
Delta R.T.: -0.009 min
Response: 551715
Conc: 21.16 ng/ml