

Method Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\
 Method File : FB091621.M
 Title :
 Last Update : Thu Sep 16 16:23:30 2021
 Response Via : Initial Calibration

Calibration Files

= S1 =FB026220.D S2 =FB026221.D
 S3 =FB026222.D S4 =FB026223.D S5 =FB026224.D

	Compound	S1	S2	S3	S4	S5	Avg		%RSD
1) t	2-Methylpentane	2.233	2.207	2.246	2.072	2.350	2.222	E4	4.49
2) t	2,2,4-Trimeth...	2.186	2.186	2.228	2.150	2.363	2.223	E4	3.75
3) t	n-Heptane	2.532	2.518	2.548	2.476	2.839	2.583	E4	5.64
4) t	Benzene	2.490	2.688	2.734	2.630	3.076	2.724	E4	7.99
5) s	AAA-TFT	1.674	1.696	1.711	1.646	1.789	1.703	E4	3.15
6) t	Toluene	2.539	2.576	2.625	2.544	2.820	2.621	E4	4.45
7) t	Ethylbenzene	2.238	2.275	2.275	2.227	2.386	2.280	E4	2.76
8) t	m-Xylene	2.228	2.252	2.243	2.181	2.305	2.242	E4	2.00
9) t	O-Xylene	2.145	2.142	2.135	2.064	2.130	2.123	E4	1.58
10) t	1,2,4-Trimeth...	2.286	2.289	2.247	2.182	2.270	2.255	E4	1.95

(#) = Out of Range