

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092018\
 Data File : FB017833.D
 Signal(s) : FID2B.CH
 Acq On : 20 Sep 2018 21:50
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Sep 21 07:51:44 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
 Quant Title :
 QLast Update : Fri Sep 14 13:09:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.106	193164	22.836 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.959	479347	30.100 ng/ml
2) t 2,2,4-Trimethylpentane	7.723	486262	30.556 ng/ml
3) t n-Heptane	8.062	155230	10.073 ng/ml
4) t Benzene	8.202	200559	10.535 ng/ml
6) t Toluene	10.955	488476	30.865 ng/ml
7) t Ethylbenzene	13.405	165348	10.119 ng/ml
8) t m-Xylene	13.537	350864	20.281 ng/ml
9) t O-Xylene	14.269	370596	20.394 ng/ml
10) t 1,2,4-Trimethylbenzene	16.552	418196	19.660 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092018\
Data File : FB017833.D
Signal(s) : FID2B.CH
Acq On : 20 Sep 2018 21:50
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 21 07:51:44 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
Quant Title :
QLast Update : Fri Sep 14 13:09:25 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 479347
Conc: 30.10 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.723 min
Delta R.T.: -0.004 min
Response: 486262
Conc: 30.56 ng/ml




#3 n-Heptane

R.T.: 8.062 min
Delta R.T.: -0.003 min
Response: 155230
Conc: 10.07 ng/ml




#4 Benzene

R.T.: 8.202 min
Delta R.T.: -0.003 min
Response: 200559
Conc: 10.54 ng/ml




#5 AAA-TFT

R.T.: 9.106 min
Delta R.T.: -0.004 min
Response: 193164
Conc: 22.84 ng/ml




#6 Toluene

R.T.: 10.955 min
Delta R.T.: -0.004 min
Response: 488476
Conc: 30.87 ng/ml




#7 Ethylbenzene

R.T.: 13.405 min
Delta R.T.: -0.004 min
Response: 165348
Conc: 10.12 ng/ml



#8 m-Xylene

R.T.: 13.537 min
Delta R.T.: -0.004 min
Response: 350864
Conc: 20.28 ng/ml




#9 O-Xylene

R.T.: 14.269 min

Delta R.T.: -0.004 min

Response: 370596

Conc: 20.39 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.552 min

Delta R.T.: -0.004 min

Response: 418196

Conc: 19.66 ng/ml