

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092022\
 Data File : FB028328.D
 Signal(s) : FID2B.CH
 Acq On : 20 Sep 2022 17:44
 Operator : AJ/MA
 Sample : N4702-01MSD 50X
 Misc : 5.08G/5.00 ML MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BU-03-091622MSD

Integration File: Calibration.e
 Quant Time: Sep 21 05:51:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090822.M
 Quant Title :
 QLast Update : Thu Sep 08 15:28:37 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.801	300415	25.100 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.707	419794	20.734 ng/ml
2) t 2,2,4-Trimethylpentane	7.439	548346	26.067 ng/ml
3) t n-Heptane	7.759	181548	8.445 ng/ml
4) t Benzene	7.896	261292	10.621 ng/ml
6) t Toluene	10.631	718595	30.866 ng/ml
7) t Ethylbenzene	13.067	227158	10.215 ng/ml
8) t m-Xylene	13.200	440112	20.168 ng/ml
9) t O-Xylene	13.927	419073	20.002 ng/ml
10) t 1,2,4-Trimethylbenzene	16.202	365257	18.166 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092022\
Data File : FB028328.D
Signal(s) : FID2B.CH
Acq On : 20 Sep 2022 17:44
Operator : AJ/MA
Sample : N4702-01MSD 50X
Misc : 5.08G/5.00 ML MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BU-03-091622MSD

Integration File: Calibration.e
Quant Time: Sep 21 05:51:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090822.M
Quant Title :
QLast Update : Thu Sep 08 15:28:37 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

