

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092220\
 Data File : FB024916.D
 Signal(s) : FID2B.CH
 Acq On : 22 Sep 2020 9:37
 Operator : DD/AJ
 Sample : BSF0922S1
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Sep 23 01:14:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091520.M
 Quant Title :
 QLast Update : Tue Sep 15 13:37:27 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	8.799	457838	20.576 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.689	923596	31.404 ng/ml
2) t	2,2,4-Trimethylpentane	7.418	969207	32.125 ng/ml
3) t	n-Heptane	7.758	319569	10.549 ng/ml
4) t	Benzene	7.892	381652	9.861 ng/ml
6) t	Toluene	10.630	1079954	29.641 ng/ml
7) t	Ethylbenzene	13.069	321853	9.575 ng/ml
8) t	m-Xylene	13.202	642724	19.620 ng/ml
9) t	O-Xylene	13.930	615829	19.280 ng/ml
10) t	1,2,4-Trimethylbenzene	16.207	515809	19.780 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092220\
Data File : FB024916.D
Signal(s) : FID2B.CH
Acq On : 22 Sep 2020 9:37
Operator : DD/AJ
Sample : BSF0922S1
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 23 01:14:28 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091520.M
Quant Title :
QLast Update : Tue Sep 15 13:37:27 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

