

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092320\
 Data File : FB024928.D
 Signal(s) : FID2B.CH
 Acq On : 23 Sep 2020 11:32
 Operator : DD/AJ
 Sample : FB092320GROICV
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 FB092320GROICV

Integration File: SAMPLE.e
 Quant Time: Sep 23 11:47:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092320.M
 Quant Title :
 QLast Update : Wed Sep 23 11:20:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.801	425146	20.251 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	892909	29.691 ng/ml
2) t 2,2,4-Trimethylpentane	7.418	921873	30.162 ng/ml
3) t n-Heptane	7.759	306570	10.077 ng/ml
4) t Benzene	7.894	374978	10.066 ng/ml
6) t Toluene	10.632	1056485	30.263 ng/ml
7) t Ethylbenzene	13.072	317129	9.783 ng/ml
8) t m-Xylene	13.205	636766	20.045 ng/ml
9) t O-Xylene	13.932	610343	20.085 ng/ml
10) t 1,2,4-Trimethylbenzene	16.210	508597	20.053 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092320\
Data File : FB024928.D
Signal(s) : FID2B.CH
Acq On : 23 Sep 2020 11:32
Operator : DD/AJ
Sample : FB092320GROICV
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
FB092320GROICV

Integration File: SAMPLE.e
Quant Time: Sep 23 11:47:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092320.M
Quant Title :
QLast Update : Wed Sep 23 11:20:20 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 892909
Conc: 29.69 ng/ml

Instrument :
FID_B
ClientSampleId :
FB092320GROICV




#2 2,2,4-Trimethylpentane

R.T.: 7.418 min
Delta R.T.: 0.000 min
Response: 921873
Conc: 30.16 ng/ml




#3 n-Heptane

R.T.: 7.759 min
Delta R.T.: 0.000 min
Response: 306570
Conc: 10.08 ng/ml



#4 Benzene


R.T.: 7.894 min
Delta R.T.: -0.001 min
Response: 374978
Conc: 10.07 ng/ml



#5 AAA-TFT


R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 425146
Conc: 20.25 ng/ml

Instrument :
FID_B
ClientSampleId :
FB092320GROICV




#6 Toluene

R.T.: 10.632 min
Delta R.T.: 0.000 min
Response: 1056485
Conc: 30.26 ng/ml




#7 Ethylbenzene

R.T.: 13.072 min
Delta R.T.: 0.000 min
Response: 317129
Conc: 9.78 ng/ml



#8 m-Xylene

R.T.: 13.205 min
Delta R.T.: 0.000 min
Response: 636766
Conc: 20.05 ng/ml



#9 O-Xylene


R.T.: 13.932 min

Delta R.T.: 0.000 min

Response: 610343

Conc: 20.09 ng/ml

Instrument :
FID_B
ClientSampleId :
FB092320GROICV



#10 1,2,4-Trimethylbenzene

R.T.: 16.210 min

Delta R.T.: 0.000 min

Response: 508597

Conc: 20.05 ng/ml