

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092518\
 Data File : FB017969.D
 Signal(s) : FID2B.CH
 Acq On : 26 Sep 2018 14:00
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Sep 26 14:33:02 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
 Quant Title :
 QLast Update : Fri Sep 14 13:09:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.111	173440	20.504 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.963	427679	26.855 ng/ml
2) t 2,2,4-Trimethylpentane	7.727	440421	27.676 ng/ml
3) t n-Heptane	8.066	136772	8.875 ng/ml
4) t Benzene	8.207	177380	9.318 ng/ml
6) t Toluene	10.960	465097	29.388 ng/ml
7) t Ethylbenzene	13.409	155285	9.503 ng/ml
8) t m-Xylene	13.542	329326	19.036 ng/ml
9) t O-Xylene	14.274	345689	19.023 ng/ml
10) t 1,2,4-Trimethylbenzene	16.556	392608	18.457 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092518\
Data File : FB017969.D
Signal(s) : FID2B.CH
Acq On : 26 Sep 2018 14:00
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 45 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 26 14:33:02 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
Quant Title :
QLast Update : Fri Sep 14 13:09:25 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.963 min
Delta R.T.: 0.003 min
Response: 427679
Conc: 26.86 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.727 min
Delta R.T.: 0.000 min
Response: 440421
Conc: 27.68 ng/ml




#3 n-Heptane

R.T.: 8.066 min
Delta R.T.: 0.001 min
Response: 136772
Conc: 8.88 ng/ml




#4 Benzene

R.T.: 8.207 min
Delta R.T.: 0.002 min
Response: 177380
Conc: 9.32 ng/ml




#5 AAA-TFT

R.T.: 9.111 min
Delta R.T.: 0.000 min
Response: 173440
Conc: 20.50 ng/ml




#6 Toluene

R.T.: 10.960 min
Delta R.T.: 0.000 min
Response: 465097
Conc: 29.39 ng/ml




#7 Ethylbenzene

R.T.: 13.409 min
Delta R.T.: 0.000 min
Response: 155285
Conc: 9.50 ng/ml




#8 m-Xylene

R.T.: 13.542 min
Delta R.T.: 0.000 min
Response: 329326
Conc: 19.04 ng/ml



#9 O-Xylene

R.T.: 14.274 min
Delta R.T.: 0.000 min
Response: 345689
Conc: 19.02 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.556 min
Delta R.T.: 0.000 min
Response: 392608
Conc: 18.46 ng/ml