

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092520\
 Data File : FB024952.D
 Signal(s) : FID2B.CH
 Acq On : 25 Sep 2020 15:00
 Operator : DD/AJ
 Sample : L4136-03MSD
 Misc : 5.05/5.00ml DI WATER
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Sep 26 00:06:55 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092320.M
 Quant Title :
 QLast Update : Wed Sep 23 11:20:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.800	391995	18.672 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	702302	23.353 ng/ml
2) t 2,2,4-Trimethylpentane	7.419	549395	17.975 ng/ml
3) t n-Heptane	7.757	176236	5.793 ng/ml
4) t Benzene	7.893	334089	8.968 ng/ml
6) t Toluene	10.631	884979	25.351 ng/ml
7) t Ethylbenzene	13.071	255601	7.885 ng/ml
8) t m-Xylene	13.204	509599	16.042 ng/ml
9) t O-Xylene	13.931	500752	16.479 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	389290	15.349 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092520\
Data File : FB024952.D
Signal(s) : FID2B.CH
Acq On : 25 Sep 2020 15:00
Operator : DD/AJ
Sample : L4136-03MSD
Misc : 5.05/5.00ml DI WATER
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 26 00:06:55 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092320.M
Quant Title :
QLast Update : Wed Sep 23 11:20:20 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

