

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092719\
 Data File : FB022623.D
 Signal(s) : FID2B.CH
 Acq On : 27 Sep 2019 12:13
 Operator : DD/AJ
 Sample : K5064-01MS
 Misc : 5.05g/5.00ML DI WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Sep 27 23:07:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
 Quant Title :
 QLast Update : Wed Sep 04 17:01:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.965	567919	19.525 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.831	1025560	20.873 ng/ml
2) t 2,2,4-Trimethylpentane	7.583	1176906	22.792 ng/ml
3) t n-Heptane	7.923	375090	7.453 ng/ml
4) t Benzene	8.058	478121	8.056 ng/ml
6) t Toluene	10.806	1409134	29.620 ng/ml
7) t Ethylbenzene	13.253	444389	14.823 ng/ml
8) t m-Xylene	13.385	877568	33.033 ng/ml
9) t O-Xylene	14.115	836237	34.470 ng/ml
10) t 1,2,4-Trimethylbenzene	16.396	564339	21.646 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092719\
Data File : FB022623.D
Signal(s) : FID2B.CH
Acq On : 27 Sep 2019 12:13
Operator : DD/AJ
Sample : K5064-01MS
Misc : 5.05g/5.00ML DI WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Sep 27 23:07:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
Quant Title :
QLast Update : Wed Sep 04 17:01:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.831 min
Delta R.T.: -0.016 min
Response: 1025560
Conc: 20.87 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.583 min
Delta R.T.: -0.014 min
Response: 1176906
Conc: 22.79 ng/ml




#3 n-Heptane

R.T.: 7.923 min
Delta R.T.: -0.015 min
Response: 375090
Conc: 7.45 ng/ml



#4 Benzene


R.T.: 8.058 min
Delta R.T.: -0.015 min
Response: 478121
Conc: 8.06 ng/ml



#5 AAA-TFT


R.T.: 8.965 min
Delta R.T.: -0.014 min
Response: 567919
Conc: 19.52 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.806 min
Delta R.T.: -0.013 min
Response: 1409134
Conc: 29.62 ng/ml




#7 Ethylbenzene

R.T.: 13.253 min
Delta R.T.: -0.012 min
Response: 444389
Conc: 14.82 ng/ml



#8 m-Xylene


R.T.: 13.385 min
Delta R.T.: -0.012 min
Response: 877568
Conc: 33.03 ng/ml



#9 O-Xylene

R.T.: 14.115 min
Delta R.T.: -0.013 min
Response: 836237
Conc: 34.47 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.396 min
Delta R.T.: -0.011 min
Response: 564339
Conc: 21.65 ng/ml