

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092818\
Data File : FB018013.D
Signal(s) : FID2B.CH
Acq On : 28 Sep 2018 19:34
Operator : MJ/SJ
Sample : J4777-39
Misc : 5.01g/5.00mL DI Water
ALS Vial : 21 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 01 07:53:30 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
Quant Title :
QLast Update : Fri Sep 14 13:09:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.108	133647	15.800 ng/ml
Target Compounds			
7) t Ethylbenzene	13.386	3568	0.218 ng/ml
8) t m-Xylene	13.537	14371	0.831 ng/ml
9) t O-Xylene	14.283	7391	0.407 ng/ml
10) t 1,2,4-Trimethylbenzene	16.555	43154	2.029 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB092818\
Data File : FB018013.D
Signal(s) : FID2B.CH
Acq On : 28 Sep 2018 19:34
Operator : MJ/SJ
Sample : J4777-39
Misc : 5.01g/5.00mL DI Water
ALS Vial : 21 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 01 07:53:30 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
Quant Title :
QLast Update : Fri Sep 14 13:09:25 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT

R.T.: 9.108 min
Delta R.T.: -0.002 min
Response: 133647
Conc: 15.80 ng/ml




#7 Ethylbenzene

R.T.: 13.386 min
Delta R.T.: -0.023 min
Response: 3568
Conc: 0.22 ng/ml




#8 m-Xylene

R.T.: 13.537 min
Delta R.T.: -0.004 min
Response: 14371
Conc: 0.83 ng/ml



#9 O-Xylene

R.T.: 14.283 min
Delta R.T.: 0.010 min
Response: 7391
Conc: 0.41 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.555 min
Delta R.T.: 0.000 min
Response: 43154
Conc: 2.03 ng/ml