

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB093019\
 Data File : FB022653.D
 Signal(s) : FID2B.CH
 Acq On : 1 Oct 2019 00:15
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Oct 01 01:26:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
 Quant Title :
 QLast Update : Wed Sep 04 17:01:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.966	540473	18.581 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.831	966148	19.664 ng/ml
2) t 2,2,4-Trimethylpentane	7.583	1091337	21.134 ng/ml
3) t n-Heptane	7.923	349137	6.937 ng/ml
4) t Benzene	8.058	457256	7.704 ng/ml
6) t Toluene	10.807	1356915	28.523 ng/ml
7) t Ethylbenzene	13.253	430236	14.351 ng/ml
8) t m-Xylene	13.386	857079	32.262 ng/ml
9) t O-Xylene	14.116	816530	33.657 ng/ml
10) t 1,2,4-Trimethylbenzene	16.396	574112	22.021 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB093019\
Data File : FB022653.D
Signal(s) : FID2B.CH
Acq On : 1 Oct 2019 00:15
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Oct 01 01:26:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB090419.M
Quant Title :
QLast Update : Wed Sep 04 17:01:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.831 min
Delta R.T.: -0.016 min
Response: 966148
Conc: 19.66 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.583 min
Delta R.T.: -0.015 min
Response: 1091337
Conc: 21.13 ng/ml




#3 n-Heptane

R.T.: 7.923 min
Delta R.T.: -0.014 min
Response: 349137
Conc: 6.94 ng/ml



#4 Benzene


R.T.: 8.058 min
Delta R.T.: -0.014 min
Response: 457256
Conc: 7.70 ng/ml



#5 AAA-TFT


R.T.: 8.966 min
Delta R.T.: -0.013 min
Response: 540473
Conc: 18.58 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.807 min
Delta R.T.: -0.013 min
Response: 1356915
Conc: 28.52 ng/ml




#7 Ethylbenzene

R.T.: 13.253 min
Delta R.T.: -0.012 min
Response: 430236
Conc: 14.35 ng/ml



#8 m-Xylene


R.T.: 13.386 min
Delta R.T.: -0.012 min
Response: 857079
Conc: 32.26 ng/ml



#9 O-Xylene

R.T.: 14.116 min
Delta R.T.: -0.012 min
Response: 816530
Conc: 33.66 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.396 min
Delta R.T.: -0.011 min
Response: 574112
Conc: 22.02 ng/ml