

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB093020\
 Data File : FB024974.D
 Signal(s) : FID2B.CH
 Acq On : 30 Sep 2020 11:01
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 01 01:00:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092320.M
 Quant Title :
 QLast Update : Wed Sep 23 11:20:20 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	393494	18.743 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.689	849103	28.234 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	902300	29.522 ng/ml
3) t n-Heptane	7.755	297756	9.788 ng/ml
4) t Benzene	7.891	357617	9.600 ng/ml
6) t Toluene	10.628	1013920	29.044 ng/ml
7) t Ethylbenzene	13.069	312810	9.650 ng/ml
8) t m-Xylene	13.201	619606	19.505 ng/ml
9) t O-Xylene	13.929	604898	19.906 ng/ml
10) t 1,2,4-Trimethylbenzene	16.207	501297	19.765 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB093020\
Data File : FB024974.D
Signal(s) : FID2B.CH
Acq On : 30 Sep 2020 11:01
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 01 01:00:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092320.M
Quant Title :
QLast Update : Wed Sep 23 11:20:20 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.689 min
Delta R.T.: -0.001 min
Response: 849103
Conc: 28.23 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.413 min
Delta R.T.: -0.005 min
Response: 902300
Conc: 29.52 ng/ml




#3 n-Heptane

R.T.: 7.755 min
Delta R.T.: -0.004 min
Response: 297756
Conc: 9.79 ng/ml




#4 Benzene

R.T.: 7.891 min
Delta R.T.: -0.004 min
Response: 357617
Conc: 9.60 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: -0.004 min
Response: 393494
Conc: 18.74 ng/ml




#6 Toluene

R.T.: 10.628 min
Delta R.T.: -0.003 min
Response: 1013920
Conc: 29.04 ng/ml




#7 Ethylbenzene

R.T.: 13.069 min
Delta R.T.: -0.003 min
Response: 312810
Conc: 9.65 ng/ml




#8 m-Xylene

R.T.: 13.201 min
Delta R.T.: -0.003 min
Response: 619606
Conc: 19.51 ng/ml



#9 O-Xylene

R.T.: 13.929 min
Delta R.T.: -0.003 min
Response: 604898
Conc: 19.91 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.207 min
Delta R.T.: -0.003 min
Response: 501297
Conc: 19.76 ng/ml