

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100118\
 Data File : FB018047.D
 Signal(s) : FID2B.CH
 Acq On : 2 Oct 2018 00:34
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 02 07:24:47 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
 Quant Title :
 QLast Update : Fri Sep 14 13:09:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

	Compound	R.T.	Response	Conc Units

System Monitoring Compounds				
5) s	AAA-TFT	9.107	182767	21.607 ng/ml
Target Compounds				
1) t	2-Methylpentane	4.960	488573	30.679 ng/ml
2) t	2,2,4-Trimethylpentane	7.725	490345	30.813 ng/ml
3) t	n-Heptane	8.064	154170	10.004 ng/ml
4) t	Benzene	8.202	194274	10.205 ng/ml
6) t	Toluene	10.957	499615	31.569 ng/ml
7) t	Ethylbenzene	13.407	165308	10.117 ng/ml
8) t	m-Xylene	13.539	351431	20.314 ng/ml
9) t	O-Xylene	14.270	370562	20.392 ng/ml
10) t	1,2,4-Trimethylbenzene	16.553	415536	19.535 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100118\
Data File : FB018047.D
Signal(s) : FID2B.CH
Acq On : 2 Oct 2018 00:34
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 31 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 02 07:24:47 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
Quant Title :
QLast Update : Fri Sep 14 13:09:25 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 488573
Conc: 30.68 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.725 min
Delta R.T.: -0.002 min
Response: 490345
Conc: 30.81 ng/ml




#3 n-Heptane

R.T.: 8.064 min
Delta R.T.: -0.001 min
Response: 154170
Conc: 10.00 ng/ml




#4 Benzene

R.T.: 8.202 min
Delta R.T.: -0.002 min
Response: 194274
Conc: 10.21 ng/ml




#5 AAA-TFT

R.T.: 9.107 min
Delta R.T.: -0.003 min
Response: 182767
Conc: 21.61 ng/ml




#6 Toluene

R.T.: 10.957 min
Delta R.T.: -0.003 min
Response: 499615
Conc: 31.57 ng/ml




#7 Ethylbenzene

R.T.: 13.407 min
Delta R.T.: -0.002 min
Response: 165308
Conc: 10.12 ng/ml




#8 m-Xylene

R.T.: 13.539 min
Delta R.T.: -0.002 min
Response: 351431
Conc: 20.31 ng/ml



#9 O-Xylene

R.T.: 14.270 min
Delta R.T.: -0.003 min
Response: 370562
Conc: 20.39 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.553 min
Delta R.T.: -0.002 min
Response: 415536
Conc: 19.53 ng/ml