

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100119\
 Data File : FB022659.D
 Signal(s) : FID2B.CH
 Acq On : 1 Oct 2019 23:01
 Operator : DD/AJ
 Sample : FB100119GROICV
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 02 06:19:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100119.M
 Quant Title :
 QLast Update : Wed Oct 02 06:15:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.970	486351	20.185 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.834	1018259	30.534 ng/ml
2) t 2,2,4-Trimethylpentane	7.588	1039306	29.963 ng/ml
3) t n-Heptane	7.927	342880	10.126 ng/ml
4) t Benzene	8.062	428430	10.287 ng/ml
6) t Toluene	10.811	1247531	30.820 ng/ml
7) t Ethylbenzene	13.256	402015	10.388 ng/ml
8) t m-Xylene	13.390	783507	21.130 ng/ml
9) t O-Xylene	14.120	743294	21.289 ng/ml
10) t 1,2,4-Trimethylbenzene	16.400	515693	20.813 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100119\
Data File : FB022659.D
Signal(s) : FID2B.CH
Acq On : 1 Oct 2019 23:01
Operator : DD/AJ
Sample : FB100119GR0ICV
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 02 06:19:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100119.M
Quant Title :
QLast Update : Wed Oct 02 06:15:33 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.834 min
Delta R.T.: 0.004 min
Response: 1018259
Conc: 30.53 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.588 min
Delta R.T.: 0.002 min
Response: 1039306
Conc: 29.96 ng/ml




#3 n-Heptane

R.T.: 7.927 min
Delta R.T.: 0.002 min
Response: 342880
Conc: 10.13 ng/ml




#4 Benzene

R.T.: 8.062 min
Delta R.T.: 0.001 min
Response: 428430
Conc: 10.29 ng/ml




#5 AAA-TFT

R.T.: 8.970 min
Delta R.T.: 0.002 min
Response: 486351
Conc: 20.19 ng/ml




#6 Toluene

R.T.: 10.811 min
Delta R.T.: 0.000 min
Response: 1247531
Conc: 30.82 ng/ml




#7 Ethylbenzene

R.T.: 13.256 min
Delta R.T.: 0.000 min
Response: 402015
Conc: 10.39 ng/ml




#8 m-Xylene

R.T.: 13.390 min
Delta R.T.: 0.000 min
Response: 783507
Conc: 21.13 ng/ml



#9 O-Xylene

R.T.: 14.120 min
Delta R.T.: 0.000 min
Response: 743294
Conc: 21.29 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.400 min
Delta R.T.: 0.000 min
Response: 515693
Conc: 20.81 ng/ml