

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100119\
 Data File : FB022656.D
 Signal(s) : FID2B.CH
 Acq On : 1 Oct 2019 21:31
 Operator : DD/AJ
 Sample : 20 GRO STD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Oct 02 06:11:08 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100119.M
 Quant Title :
 QLast Update : Wed Oct 02 06:09:38 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.969	447162	20.000 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.833	914715	30.000 ng/ml
2) t 2,2,4-Trimethylpentane	7.587	957336	30.000 ng/ml
3) t n-Heptane	7.927	310941	10.000 ng/ml
4) t Benzene	8.061	394549	10.000 ng/ml
6) t Toluene	10.810	1158456	30.000 ng/ml
7) t Ethylbenzene	13.256	374235	10.000 ng/ml
8) t m-Xylene	13.390	727710	20.000 ng/ml
9) t O-Xylene	14.119	690784	20.000 ng/ml
10) t 1,2,4-Trimethylbenzene	16.401	493891	20.000 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100119\
Data File : FB022656.D
Signal(s) : FID2B.CH
Acq On : 1 Oct 2019 21:31
Operator : DD/AJ
Sample : 20 GRO STD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Oct 02 06:11:08 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100119.M
Quant Title :
QLast Update : Wed Oct 02 06:09:38 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.833 min
Delta R.T.: 0.000 min
Response: 914715
Conc: 30.00 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.587 min
Delta R.T.: 0.000 min
Response: 957336
Conc: 30.00 ng/ml




#3 n-Heptane

R.T.: 7.927 min
Delta R.T.: 0.000 min
Response: 310941
Conc: 10.00 ng/ml




#4 Benzene

R.T.: 8.061 min
Delta R.T.: 0.000 min
Response: 394549
Conc: 10.00 ng/ml




#5 AAA-TFT

R.T.: 8.969 min
Delta R.T.: 0.000 min
Response: 447162
Conc: 20.00 ng/ml




#6 Toluene

R.T.: 10.810 min
Delta R.T.: 0.000 min
Response: 1158456
Conc: 30.00 ng/ml




#7 Ethylbenzene

R.T.: 13.256 min
Delta R.T.: 0.000 min
Response: 374235
Conc: 10.00 ng/ml




#8 m-Xylene

R.T.: 13.390 min
Delta R.T.: 0.000 min
Response: 727710
Conc: 20.00 ng/ml



#9 O-Xylene

R.T.: 14.119 min
Delta R.T.: 0.000 min
Response: 690784
Conc: 20.00 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.401 min
Delta R.T.: 0.000 min
Response: 493891
Conc: 20.00 ng/ml