

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100219\
 Data File : FB022669.D
 Signal(s) : FID2B.CH
 Acq On : 2 Oct 2019 17:42
 Operator : DD/AJ
 Sample : K5160-01MSD
 Misc : 5.03g/5.00ml DI WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 CSA-3-1MSD

Integration File: Calibration.e
 Quant Time: Oct 03 02:04:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100119.M
 Quant Title :
 QLast Update : Wed Oct 02 06:15:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.973	457119	18.972 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.837	871851	26.144 ng/ml
2) t 2,2,4-Trimethylpentane	7.590	936025	26.985 ng/ml
3) t n-Heptane	7.930	301344	8.899 ng/ml
4) t Benzene	8.064	387331	9.300 ng/ml
6) t Toluene	10.813	1110050	27.424 ng/ml
7) t Ethylbenzene	13.259	354815	9.168 ng/ml
8) t m-Xylene	13.392	682971	18.419 ng/ml
9) t O-Xylene	14.122	652337	18.684 ng/ml
10) t 1,2,4-Trimethylbenzene	16.403	467299	18.860 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

```
Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100219\
Data File : FB022669.D
Signal(s) : FID2B.CH
Acq On    : 2 Oct 2019 17:42
Operator  : DD/AJ
Sample    : K5160-01MSD
Misc      : 5.03g/5.00ml DI WATER
ALS Vial  : 10 Sample Multiplier: 1
```