

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100419\
 Data File : FB022681.D
 Signal(s) : FID2B.CH
 Acq On : 4 Oct 2019 8:19
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 04 21:38:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100119.M
 Quant Title :
 QLast Update : Wed Oct 02 06:15:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.960	502933	20.874 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.830	1042806	31.270 ng/ml
2) t 2,2,4-Trimethylpentane	7.579	1072493	30.920 ng/ml
3) t n-Heptane	7.918	360675	10.651 ng/ml
4) t Benzene	8.053	426419	10.239 ng/ml
6) t Toluene	10.801	1244588	30.747 ng/ml
7) t Ethylbenzene	13.248	398223	10.290 ng/ml
8) t m-Xylene	13.381	743526	20.052 ng/ml
9) t O-Xylene	14.110	696607	19.952 ng/ml
10) t 1,2,4-Trimethylbenzene	16.392	537200	21.681 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100419\
Data File : FB022681.D
Signal(s) : FID2B.CH
Acq On : 4 Oct 2019 8:19
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 04 21:38:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100119.M
Quant Title :
QLast Update : Wed Oct 02 06:15:33 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.830 min
Delta R.T.: 0.000 min
Response: 1042806
Conc: 31.27 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.579 min
Delta R.T.: -0.007 min
Response: 1072493
Conc: 30.92 ng/ml




#3 n-Heptane

R.T.: 7.918 min
Delta R.T.: -0.008 min
Response: 360675
Conc: 10.65 ng/ml




#4 Benzene

R.T.: 8.053 min
Delta R.T.: -0.008 min
Response: 426419
Conc: 10.24 ng/ml




#5 AAA-TFT

R.T.: 8.960 min
Delta R.T.: -0.008 min
Response: 502933
Conc: 20.87 ng/ml




#6 Toluene

R.T.: 10.801 min
Delta R.T.: -0.009 min
Response: 1244588
Conc: 30.75 ng/ml




#7 Ethylbenzene

R.T.: 13.248 min
Delta R.T.: -0.008 min
Response: 398223
Conc: 10.29 ng/ml




#8 m-Xylene

R.T.: 13.381 min
Delta R.T.: -0.009 min
Response: 743526
Conc: 20.05 ng/ml



#9 O-Xylene

R.T.: 14.110 min
Delta R.T.: -0.009 min
Response: 696607
Conc: 19.95 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.392 min
Delta R.T.: -0.008 min
Response: 537200
Conc: 21.68 ng/ml