

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100424\
Data File : FB030970.D
Signal(s) : FID2B.CH
Acq On : 4 Oct 2024 13:15
Operator : YP/AJ
Sample : VBF1004S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
VBF1004S1

Integration File: Calibration.e
Quant Time: Oct 05 04:15:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100424.M
Quant Title :
QLast Update : Fri Oct 04 11:53:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	606157	20.220 ng/ml
Target Compounds			

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100424\
Data File : FB030970.D
Signal(s) : FID2B.CH
Acq On : 4 Oct 2024 13:15
Operator : YP/AJ
Sample : VBF1004S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
VBF1004S1

Integration File: Calibration.e
Quant Time: Oct 05 04:15:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100424.M
Quant Title :
QLast Update : Fri Oct 04 11:53:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um





#5 AAA-TFT

R.T.: 8.791 min
Delta R.T.: -0.005 min
Response: 606157
Conc: 20.22 ng/ml

Instrument :
FID_B
ClientSampleId :
VBF1004S1