

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100518\
 Data File : FB018119.D
 Signal(s) : FID2B.CH
 Acq On : 5 Oct 2018 9:57
 Operator : MJ/SJ
 Sample : J5194-01
 Misc : 5.02g/5.00mL DI Water
 ALS Vial : 5 Sample Multiplier: 1

Integration File: SAMPLE.e
 Quant Time: Oct 05 10:19:15 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
 Quant Title :
 QLast Update : Fri Sep 14 13:09:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.109	124715	14.744 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.977	4059	0.255 ng/ml
2) t 2,2,4-Trimethylpentane	7.717	1360	0.085 ng/ml
3) t n-Heptane	8.069	2185	0.142 ng/ml
4) t Benzene	8.206	1180	0.062 ng/ml
6) t Toluene	10.964	360	0.023 ng/ml
7) t Ethylbenzene	13.370	96106	5.882 ng/ml
9) t O-Xylene	14.275	29185	1.606 ng/ml
10) t 1,2,4-Trimethylbenzene	16.587	12596	0.592 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100518\
Data File : FB018119.D
Signal(s) : FID2B.CH
Acq On : 5 Oct 2018 9:57
Operator : MJ/SJ
Sample : J5194-01
Misc : 5.02g/5.00mL DI Water
ALS Vial : 5 Sample Multiplier: 1

Integration File: SAMPLE.e
Quant Time: Oct 05 10:19:15 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
Quant Title :
QLast Update : Fri Sep 14 13:09:25 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.977 min
Delta R.T.: 0.017 min
Response: 4059
Conc: 0.25 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.717 min
Delta R.T.: -0.010 min
Response: 1360
Conc: 0.09 ng/ml




#3 n-Heptane

R.T.: 8.069 min
Delta R.T.: 0.004 min
Response: 2185
Conc: 0.14 ng/ml




#4 Benzene

R.T.: 8.206 min
Delta R.T.: 0.002 min
Response: 1180
Conc: 0.06 ng/ml




#5 AAA-TFT

R.T.: 9.109 min
Delta R.T.: -0.001 min
Response: 124715
Conc: 14.74 ng/ml




#6 Toluene

R.T.: 10.964 min
Delta R.T.: 0.005 min
Response: 360
Conc: 0.02 ng/ml




#7 Ethylbenzene

R.T.: 13.370 min
Delta R.T.: -0.039 min
Response: 96106
Conc: 5.88 ng/ml



#9 O-Xylene

R.T.: 14.275 min
Delta R.T.: 0.002 min
Response: 29185
Conc: 1.61 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.587 min
Delta R.T.: 0.032 min
Response: 12596
Conc: 0.59 ng/ml