

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101018\
 Data File : FB018168.D
 Signal(s) : FID2B.CH
 Acq On : 10 Oct 2018 15:47
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Oct 10 16:06:52 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
 Quant Title :
 QLast Update : Fri Sep 14 13:09:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.111	162542	19.216 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.961	417207	26.198 ng/ml
2) t 2,2,4-Trimethylpentane	7.726	439220	27.600 ng/ml
3) t n-Heptane	8.067	136782	8.876 ng/ml
4) t Benzene	8.206	185962	9.769 ng/ml
6) t Toluene	10.960	454495	28.718 ng/ml
7) t Ethylbenzene	13.409	159685	9.773 ng/ml
8) t m-Xylene	13.542	337047	19.483 ng/ml
9) t O-Xylene	14.274	350517	19.289 ng/ml
10) t 1,2,4-Trimethylbenzene	16.556	412991	19.415 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101018\
Data File : FB018168.D
Signal(s) : FID2B.CH
Acq On : 10 Oct 2018 15:47
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Oct 10 16:06:52 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB091418.M
Quant Title :
QLast Update : Fri Sep 14 13:09:25 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.961 min
Delta R.T.: 0.001 min
Response: 417207
Conc: 26.20 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.726 min
Delta R.T.: 0.000 min
Response: 439220
Conc: 27.60 ng/ml




#3 n-Heptane

R.T.: 8.067 min
Delta R.T.: 0.002 min
Response: 136782
Conc: 8.88 ng/ml




#4 Benzene

R.T.: 8.206 min
Delta R.T.: 0.002 min
Response: 185962
Conc: 9.77 ng/ml




#5 AAA-TFT

R.T.: 9.111 min
Delta R.T.: 0.001 min
Response: 162542
Conc: 19.22 ng/ml




#6 Toluene

R.T.: 10.960 min
Delta R.T.: 0.000 min
Response: 454495
Conc: 28.72 ng/ml




#7 Ethylbenzene

R.T.: 13.409 min
Delta R.T.: 0.000 min
Response: 159685
Conc: 9.77 ng/ml




#8 m-Xylene

R.T.: 13.542 min
Delta R.T.: 0.001 min
Response: 337047
Conc: 19.48 ng/ml



#9 O-Xylene

R.T.: 14.274 min
Delta R.T.: 0.000 min
Response: 350517
Conc: 19.29 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.556 min
Delta R.T.: 0.000 min
Response: 412991
Conc: 19.42 ng/ml