

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101022\
 Data File : FB028602.D
 Signal(s) : FID2B.CH
 Acq On : 11 Oct 2022 14:47
 Operator : AJ/MA
 Sample : N5018-03 50X
 Misc : 2.30G/5.00 ML DI MEOH
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Oct 12 04:49:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100522.M
 Quant Title :
 QLast Update : Wed Oct 05 12:47:25 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	289055	22.343 ng/ml
Target Compounds			
4) t Benzene	7.940	18900	0.860 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101022\
Data File : FB028602.D
Signal(s) : FID2B.CH
Acq On : 11 Oct 2022 14:47
Operator : AJ/MA
Sample : N5018-03 50X
Misc : 2.30G/5.00 ML DI MEOH
ALS Vial : 43 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Oct 12 04:49:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB100522.M
Quant Title :
QLast Update : Wed Oct 05 12:47:25 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#4 Benzene

R.T.: 7.940 min
Delta R.T.: 0.052 min
Response: 18900
Conc: 0.86 ng/ml

Instrument :
FID_B
ClientSampleId :



#5 AAA-TFT

R.T.: 8.795 min
Delta R.T.: 0.002 min
Response: 289055
Conc: 22.34 ng/ml